



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. |
|-----------------|-------------|----------------------|---------------------|------------------|
| 10/748,610 | 12/30/2003 | Robert C. Thistle | BSI-521US | 7486 |

7590 10/18/2007
Paul F. Prestia
RatnerPrestia
One Westlakes (Berwyn), Suite 301
P.O. Box 980
Valley Forge, PA 19482

| |
|----------|
| EXAMINER |
|----------|

PRONE, CHRISTOPHER D

| | |
|----------|--------------|
| ART UNIT | PAPER NUMBER |
|----------|--------------|

3738

| | |
|-----------|---------------|
| MAIL DATE | DELIVERY MODE |
|-----------|---------------|

10/18/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/748,610

Applicant(s)

THISTLE, ROBERT C.

Examiner

Christopher D. Prone

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 17 August 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-13 and 15-23 is/are pending in the application.
- 4a) Of the above claim(s) 3, 9, 17, 22 and 23 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1, 2, 4-8, 10-13, 15, 16 and 18-21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 1, 2, 4-8, 10-13, 15, 16, and 18-21 rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter, which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Each of the independent claims has been amended to recite that the fastening element "extending beyond said outer layer in no more than two locations". This recitation is not supported in the specification anywhere and the drawings directly contradict this statement. The applicant is referred to figure 4, which shows a close-up of fastening element 18. The fastening element is made up of a plurality of locations, which a great many more than two are extended outside the outer layer. A figure is attached for further clarification.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 2, 4-8, and 10-13 are rejected under 35 U.S.C. 102(b) as being anticipated by USPN 5,527,353 Schmitt.

Schmitt discloses the same invention being a nonporous graft 10. The graft comprises both an inner layer of nonporous ePTFE 4 and an outer layer 5 of woven polyester, which are laminated together. Schmitt also discloses the use of a fastening element 6 that is fixed between the inner and outer layers, which forms d-shaped hooks that may project through elements of a stent see figure 3.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 15-21 are rejected under 35 U.S.C. 103(a) as being unpatentable over USPN 5,527,353 Schmitt in view of USPN 5,891,193 Robinson.

Schmitt discloses the invention substantially as claimed being a nonporous graft 10. The graft comprises both an inner layer of nonporous ePTFE 4 and an outer layer 5 of woven polyester, which are laminated together. Schmitt also discloses the use of a fastening element 6 that is fixed between the inner and outer layers, which forms d-shaped hooks that may project through elements of a stent see figure 3. Schmitt teaches the use of the fastening elements for enhanced suture retention when suturing

Art Unit: 3738

the implant to a vessel. However Schmitt does not disclose the use of a stent attached to the graft.

Robinson teaches the use of a stent graft vascular prosthesis wherein the graft is sutured to the prosthesis in the same field of endeavor for the purpose of providing a secure connection that will keep the stent and graft connected throughout the expansion and implantation.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to combine suture the expanding stent to the graft as taught by Schmitt in order to provide a more rigid prosthesis that has enhanced expanding properties.

Response to Arguments

Applicant's arguments filed 8/17/07 have been fully considered but they are not persuasive. The applicant's amendments fail to overcome the art of record because they are deemed to be new matter for this application. In view of this new matter, the art of record is as close to the claim language as that of the current application. The art of record discloses fasteners that extend beyond the outer layer in no more than two locations as much as that of the current application.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

Art Unit: 3738

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


CDP

Christopher D Prone
Examiner
Art Unit 3738


CORRINE McDERMOTT
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700




FIG. 3




FIG. 4